


MEK-3DL
HEMATOLOGY CONTROL
2.0mL
ASSAY SHEET

LOT	B242L
	2024-05-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order.
If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420	MEK -6500 -6510	MEK -6450	MEK -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±0.6	2.0	2.0	2.0	2.0	2.02
RBC	10 ⁶ /μL	±0.15	2.41	2.41	2.41	2.41	2.36
HGB	g/dL	±0.4	6.0	6.1	6.0	6.1	6.06
HCT	%	±2.0	17.7	17.7	17.7	17.7	19.6
MCV	fL	±9.0	73.5	73.5	73.5	73.5	83.0
MCH	pg	±2.4	24.9	25.3	24.9	25.3	25.7
MCHC	g/dL	±3.0	33.9	34.5	33.9	34.5	30.9
PLT	10 ³ /μL	±30	67	67	67	67	75.0
MPV	fL	±3.0	5.8	8.2	5.8	8.2	8.1
RDW-CV *	%	±3.0	18.2	18.2	18.2	18.2	16.7
LY%	%	±10.0	68.6	68.6	68.6	68.6	68.54
MO%	%	±5.0	6.9	6.9	6.9	6.9	3.94
GR%	%	±10.0	24.5	24.5	24.5	24.5	27.52
LY	10 ³ /μL	±0.5	1.4	1.4	1.4	1.4	1.38
MO	10 ³ /μL	±0.2	0.1	0.1	0.1	0.1	0.08
GR	10 ³ /μL	±0.5	0.5	0.5	0.5	0.5	0.56

MEK-1305
QC ESR
48±17 mm
Manual Registration

Lot	242
Level	Low
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.


<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU

TOKYO 161-8560 JAPAN



MEK-3DL



B242L



2024-05-05



0694-9016840