




MK-RE2

HEMATOLOGY CONTROL
for reticulocyte

3.0mL

ASSAY SHEET

 LOT	MKRE245-2
	2024-07-05
 OPEN	14 DAYS

		MEK-9200	RANGE
RET%	%	4.15	±2.85
RET	10 ⁶ /μL	0.1488	±0.1022

2D Barcode Entry



• If RBC value is under 3.00 [$\times 10^6/\mu\text{L}$], please refrain from using.
It may have deteriorated or hemolyzed, which affects the measurement accuracy.



<https://www.nihonkohden.com/spi/index.html>

Manual Registration

Lot	245					
Level	2					
Expiration	7	M	5	D	2024	Y
Control	MKRE					

CE
0123

IVD



NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN